

DEC 10 2010

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130288✓
Name	Sgt. Nick Zito	ID#	4998✓
Date	12-10-10✓		
A	Agency	AK State Troopers	Phone # 675 4398
	Instrument Location	Aniak Post	
B	Alco S/N	56799✓	Target Value .084✓
			High Pressure 200
	Alco Test Values	.087✓ 1st Alco	.087✓ 2nd Alco
	Signature	[Signature] NPZD	
		(OVER)	

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 18th day of JAN, 2011.

Daniel L. Nicholas
Notary Public, State of Alaska
Commission Expires with Office

(Notary Seal Stamp)



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130288

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130288

DECEMBER 10, 2010

OPERATOR'S NAME:

ZITO/NICHOLAS/PAUL

OPERATOR'S NUMBER: 4998

SUBJECT'S LAST NAME:

TEST

SUBJECT'S FIRST NAME/MI :

TEST

O.L. #: TEST

DEPT/AGENCY: AST/AN10

CASE/REPORT: TEST

TEST TYPE: T

ALCO TARGET VALUE: .084

ALCO S/N: 56799

— BREATH ANALYSIS —

.084 ADJUSTED FOR 30.07 in		
ALCO TARGET	.084	10:21
BLANK TEST	.000	10:22
INTERNAL STANDARD	VERIFIED	10:22
ALCO TV 30.07 in	.087	10:23
BLANK TEST	.000	10:24
SUBJECT SAMPLE	.000	10:24
BLANK TEST	.000	10:25
ALCO TV 30.07 in	.087	10:25
BLANK TEST	.000	10:26

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130288

DECEMBER 10, 2010

TIME 10:51

— DIAGNOSTIC CHECK —

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	43c
BAROMETER:	30.07 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

!""#\$%&'()*+,-./0123456789:;<=?@ABCDEFGHIJ
KLMNOPQRSTUVWXYZ[\]^_`abcde fghijklmno
pqrstu vwxyz{|}~■